

Palintest

Water Analysis Technologies

Hydrogen Peroxide HR

Photometric test for Hydrogen Peroxide (high range) in process water and swimming pool water

COLOUR CHANGE

Colourless to Yellow

RANGE:

0 - 100 mg/L H₂O₂

TECHNICAL RESOURCE:

Hydrogen Peroxide HR Test Method, Technical Information

NOTES:

Sample needs to be free of any other oxidising agents, such as chlorine, bromine, ozone etc

01




Fill to the 10 mL line with sample.

02



Use this to "blank" the instrument. Place in cell holder and press Blank.

03



Prepare sample by adding an Acidifying PT tablet and a Hydrogen Peroxide HR tablet. Crush both tablets and stir.

04



Place in cell holder and press Measure to obtain concentration of Hydrogen Peroxide.